



**MET Laboratories, Inc.** Safety Certification - EMI - Telecom - Environmental Simulation - NEBS  
914 WEST PATAPSCO AVENUE • BALTIMORE, MARYLAND 21230-3432 • PHONE (410) 919-1802 • FAX (410) 354-3313

August 19, 2009

Mr. Kevin Moses  
Environmental Analytics Inc. DBA: LDARtools  
1320 Hwy 3 South, Unit D3  
League City, TX 77573 U.S.A.

Subject: Flame Ionization Detector, model phx21  
Listing Number E112898; MET Project Number 26721  
Safety Standards: 

- UL 61010-1 UL Standard for Safety Electrical Equipment For Measurement, Control, and Laboratory Use; Part 1: General Requirements - Second Edition
- UL913, Fifth Edition: Intrinsically Safe Apparatus and Associated Apparatus for Use in Class I, II, and III, Division 1, Hazardous (Classified) Locations
- CAN/CSA-C22.2 No. 157-92, Third Edition, Intrinsically Safe and Non-incendive Equipment for Use in Hazardous Locations

Dear Mr. Moses:

Congratulations on successfully completing the MET Certification process for the Flame Ionization Detector, model phx21. Environmental Analytics Inc. DBA: LDARtools may begin to apply the MET Mark on the above stated product at this time in accordance with the MET Mark Utilization Agreement or the MET Applicant Contract. The reports covering the above stated product will be forthcoming.

Follow-up inspections are conducted unannounced and biannually to assure the Certified product is identical to the product evaluated.

Thank you for the opportunity to perform this service for Environmental Analytics Inc. DBA: LDARtools. We look forward to future opportunities with your company.

*Sincerely,*

MET LABORATORIES, INC.

Rick Cooper  
Director of Laboratory Operations,  
Safety Laboratory

/jo



*The Nation's First Nationally Recognized Testing Laboratory*  
*MET Laboratories, Inc. is accredited by OSHA and the Standards Council of Canada.*

**NRTL**

Canadian Certification has been granted under a System 3 program as defined in ISO Guide 67.